


POC BATTERY CHART

SimplyGo



		PULSE DOSE SETTING						CONTINUOUS FLOW SETTING		
		1	2	3	4	5	6	0.5	1	2
Number of Batteries	1	3.7 hours	3.5 hours	3.1 hours	2.7 hours	2.3 hours	1.9 hours	3.1 hours	2.3 hours	.7 hours
	2	7.4 hours	7 hours	6.2 hours	5.4 hours	4.6 hours	3.8 hours	6.2 hours	4.6 hours	1.4 hour
	3	11.1 hours	10.5 hours	9.3 hours	8.1 hours	6.9 hours	5.7 hours	9.3 hours	6.9 hours	2.1 hours
	4	14.8 hours	14 hours	12.4 hours	10.8 hours	9.2 hours	7.6 hours	12.4 hours	9.2 hours	2.8 hours

 SimplyGo Mini	Battery Duration	Up to 4.5 hours (Pulse setting of 2 at 20 BPM with Standard battery) Up to 9 hours (Pulse setting of 2 at 20 BPM with Extended battery)
	* Battery Duration times are based on new, fully charged batteries used with a new SimplyGo Mini system. Battery duration times will degrade with battery age, environmental operating conditions, use over time, and operational condition of the concentrator.	

POC BATTERY CHART

Typical New Power Cartridge Operating Times

Flow Rate Setting	Continuous Flow (LPM)	mL SETTING	Pulse Dose Mode (12 BPM)
0.5	4.4 hours	-	-
1.0	3.7 hours	16mL 1.0	5.4 hours
2.0	2.0 hours	32mL 2.0	5.1 hours
3.0	1.3 hours	48mL 3.0	4.9 hours
		64mL 4.0	4.0 hours
		80mL 5.0	3.7 hours
		96mL 6.0	3.5 hours
		128mL	2.5 hours
eclipse with autoSAT®		160mL	2.0 hours
		192mL	1.7 hours